



CERTIFICATE BA-0006
NUMBER _____

CERTIFICATE OF TYPE TEST FOR OIL CONTENT METERS

THIS IS TO CERTIFY THAT THE SYSTEM COMPROMISING THE EQUIPMENT LISTED HAS BEEN EXAMINED AND TESTED IN ACCORDANCE WITH THE REQUIREMENTS OF THE SPECIFICATION CONTAINED IN PART III OF THE ANNEX TO THE RECOMMENDATION CONTAINED IN IMO RESOLUTION NO. A. 393 (X). THIS CERTIFICATE IS VALID ONLY FOR SUCH A SYSTEM.

OSM BILGE ALARM

* SYSTEM MANUFACTURED BY Babcock - Bristol Limited, Croydon, England INCLUDING

OIL CONTENT METER MANUFACTURED BY Babcock - Bristol Limited

TO DRAWING NUMBERS 101984, 101973, C012878, C012879

* ALARM MANUFACTURED BY Babcock - Bristol Limited

TO DRAWING NUMBERS C012879

CONTROL SYSTEM MANUFACTURED BY Babcock- Bristol Limited

TO DRAWING NUMBERS C012879

THE SYSTEM IS ACCEPTABLE FOR THE FOLLOWING APPLICATION (* DELETE AS APPROPRIATE)

* -- 15 PPM BILGE OR FUEL TANK BALLAST ALARM

TEST DATA AND RESULTS ATTACHED AS APPENDIX as per U.K. Dept. of Trade certificate of Type Test for OSM Bilge Alarm, dated July 18, 1980

A COPY OF THIS TEST CERTIFICATE SHOULD BE CARRIED ABOARD A VESSEL FITTED WITH THIS EQUIPMENT AT ALL TIMES



P J Strain P.J. Strain
(SIGNATURE OF AUTHORIZED INSPECTOR)

DATED THIS 13th DAY OF August 19 85



Canadian Coast Guard / Garde coterie canadienne

Canada